

Effect of surface-treatments on flexibility and guided bone regeneration of titanium barrier membrane

Jin-Tae Kim, Byoung Soo Kim, Hee Seok Jeong, Young Ku Heo, Sang-Wan Shin*, Jeong-Yol Lee*, Young Ho Shim* and Deuk Yong Lee**,[†]

Dental Materials Research Center, Neobiotech Co., Ltd., Seoul 152-582, Korea

**Institute for Clinical Dental Research, Korea University, Seoul 136-701, Korea*

***Department of Biomedical Engineering, Daelim University, Anyang 431-715, Korea*

(Received April 7, 2015)

(Revised May 6, 2015)

(Accepted May 8, 2015)

Abstract Titanium barrier membranes are prepared to investigate the effect of surface-treatments, such as machining, electropolishing, anodizing, and electropolishing + TiN coating, on the biocompatibility and physical properties of the membranes. The surface roughness (Ra) of the membrane decreases from machining ($0.37 \pm 0.09 \mu\text{m}$), TiN coating ($0.22 \pm 0.09 \mu\text{m}$), electropolishing ($0.20 \pm 0.03 \mu\text{m}$), to anodizing ($0.15 \pm 0.03 \mu\text{m}$). The highest ductility (24.50 %) is observed for the electropolished Ti membrane. No evidence of causing cell lysis or toxicity is found for the membranes regardless of the surface-treatments. Cell adhesion results of L-929 and MG-63 show that the machined Ti membrane exhibits the highest cell adhesion while the electropolished membrane is the best membrane for the L-929 cell proliferation after 7 days. However, no appreciable difference in MG-63 cell proliferation among variously surface-treated membranes is detected, suggesting that the electropolished Ti membrane is likely to be the best membrane due to the synergic combination of tailored flexibility and excellent fibroblast proliferation.

Key words Pure titanium, Surface-treatment, Roughness, Biocompatibility, Cytotoxicity, Cell proliferation

1. Introduction

Commercially pure titanium (Cp Ti) has been considered as the biocompatible metallic materials for orthopedic and dentistry implants because of its excellent mechanical properties, corrosion resistance, and biocompatibility [1-3]. The excellent biocompatibility of Ti is associated with its stable and inert oxide layer, leading to low level of electronic conductivity, high corrosion resistance, thermodynamic state at physiological pH values, and low ion-formation tendency in aqueous environment [1]. Ti has been widely used for the implant devices replacing failed hard tissue.

When tooth extraction is required it typically results in additional bone loss in buccal or lingual cortical plate. The guided bone regeneration (GBR) using the barrier membrane is useful in bone augmentation [4-6]. The GBR and the implant fixation at an infected site are frequently complicated by soft-tissue dehiscence, membrane exposure, and implant failure due to deficient

mature bone mass and quality to support the implant. The principle of the GBR is the creation and maintenance of sufficient space underneath the barrier.

The mechanical properties of the membranes should avoid collapse of the membrane and provide a constant volume underneath it. The porous areas are small enough to prevent soft tissue penetration through the membrane with no compromise in diffusion of interstitial fluid [4-6]. The non-absorbable or the resorbable barrier membranes are likely to collapse due to improper mechanical properties, resulting in insufficient space underneath the membrane for new bone to form [4]. The barriers should be left in place for up to 6 months for bone regeneration. Among the membranes, a Ti membrane with microperforation is used as the barrier because of the rigidity of Ti membrane. It makes the barrier suitable to maintain the space and cells and fluids necessary for nourishment can be passed through the perforated holes. The porous membranes are attributed to their osteoconductive properties to facilitate the migration of osteoblasts from surrounding bone into the implant site [7, 8]. However, the Ti barrier membranes need a secondary surgery to remove the barrier. In addition, it is difficult to insert the Ti membrane than flexible mem-

[†]Corresponding author
Tel: +82-31-467-4835
Fax: +82-31-467-4432
E-mail: dylee@daelim.ac.kr

branes due to the sharp edges of Ti membrane. Nevertheless, the Ti membranes have been widely used due to the biocompatibility and osteoconductivity of porous Ti membrane with controllable pore shape, pore size, and porosity. In this study, the biocompatibility and the physical properties of the Ti barrier membranes prepared by different surface treatments are studied.

2. Materials and Methods

2.1. Materials

Ti specimens with a dimension of $10(\phi) \times 0.15(t) \text{ mm}^2$ are prepared from grade 2 titanium [1]. The specimens for *in-vitro* test are then surface-treated (electropolishing, anodizing, electropolishing + TiN coating) to investigate the effect of surface roughness on the biocompatibility of the Ti membrane. The samples without any surface treatments are used as control group. Prior to the test, the specimens are γ -ray sterilized [9,10].

The electrolyte for electropolishing is prepared by mixing H_2SO_4 , H_3PO_4 , and distilled water. The specimens are then electropolished for 20–40 s at $75\sim 85^\circ\text{C}$ with a voltage of 7 V and currents of 40–50 A. Then, they are thoroughly rinsed with distilled water and neutralized with NaOH, followed by rinsing with distilled water and drying. Prior to anodizing, the specimens are cleaned by dipping in trichloroethylene for 15 min, followed by surface activation in a solution mixture of HF and HNO_3 for 30 s. The Ti specimen is then anodized for 20–30 s in potentiostatic mode (4–5 V and 3–5 A) at room temperature in the electrolytes (H_3PO_4 , H_2SO_4 , H_2O). After anodizing, the specimens are neutralized with NaOH and then dried. For TiN coating, the as-electropolished specimens are first surface cleaned by sonicating for 10 min using ethanol and acetone, followed by Ar plasma treatment for 10 min under vacuum (3×10^{-5} torr). Then, the TiN coated specimens are obtained for 5 min at 300°C in N_2 gas atmosphere (10–20 mtorr).

2.2. Surface roughness and tension test

The laser surface roughness tester (Lext OLS4100, Olympus, Japan) and the electron microscopy (CX-100S, Coxem, Korea) are employed to evaluate the roughness of the surface-treated Ti specimens. The specimens with a dimension of $2 \times 25 \times 0.15(t) \text{ mm}^3$ are surface-treated and then mechanical properties are examined by using a tension tester (E3000, Instron, UK).

2.3. Biocompatibility

2.3.1. Cell culture

MG-63 (Korean Cell Line Bank, Korea) osteoblast-like cells and L-929 (Korean Cell Line Bank, Korea) mouse fibroblast cells are used. All monolayers are incubated in a mixture of Dulbecco's minimum essential media (DMEM, Gibco-BRL, Life Tech., New Zealand) and RPMI1640 (Gibco-BRL, Life Tech., New Zealand) at 37°C in the presence of 5 % CO_2 . The DMEM is supplemented with 10 % of fetal bovine serum (Gibco-BRL, Life Tech., New Zealand), $1 \times$ antibiotic-antimycotic solution (contains 10,000 units of penicillin(base), 10,000 $\mu\text{g}/\text{ml}$ of streptomycin), and 25 g of amphotericin (Gibco-BRL, Life Tech., New Zealand) [11]. After 80 % of cell proliferation, cell detachment is done with trypsin-EDTA (0.25 % trypsin and 0.1 % glucose dissolved in 1 mM of DETA-saline, Invitrogen Co., USA), followed by subculture.

2.3.2. Cell adhesion

The 25 cm^2 culture flasks are seeded with 8×10^4 cells/ml and incubated for 3 h to evaluate the bond strength between the surface-treated specimens and the cells. The detached cells are excluded. The cells are stained by immersing in the solution of 0.2 % crystal violet (Georgiachem., Japan) dissolved in 10 % ethanol for 5 min at room temperature. After immersing the stained cells in 95 % ethanol for 15 min, the change of absorption at 595 nm is examined by using the ELISA analyzer (Spectra MAX 250s, Molecular Devices Co., USA).

2.3.3. Cell proliferation

The culture flasks are seeded with 2×10^4 cells/ml and incubated for 3 or 7 days. The specimens are added to each flask. The cell proliferation is evaluated by using the XTT assay kit (Cell Proliferation Kit II, Roche Applied Science, Germany). After incubation, XTT labeling reagent of 250 μl is added to each culture and the cultures are incubated for 2 h at 37°C in the presence of 5 % CO_2 . The change of absorption at 450 nm is examined by using the ELISA analyzer.

2.3.4. Cell cytotoxicity

The extract test method is conducted to evaluate the potential of cytotoxicity on the basis of the International Organization for Standardization (ISO 10993-5). The test samples are used within 24 h after completion of the preparation. The test extract is placed onto two




Fig. 1. 3D laser microscopic images of Ti surfaces treated by (a) machining, (b) TiN coating, (c) electropolishing, and (d) anodizing, respectively.




Fig. 2. SEM images of Ti surfaces treated by (a) machining, (b) TiN coating, (c) electropolishing, and (d) anodizing, respectively.

separate confluent monolayers of L-929 mouse fibroblast cells (1×10^5 cells/ml) propagated in 5 % CO_2 . After incubation, XTT labeling reagent of 250 μl is added to each culture and the cultures are incubated for 2 h at 37°C in the presence of 5 % CO_2 . The change of absorption at 450 nm is examined by using the ELISA analyzer. For analyses of results, statistical calculations such as t-test ($p < 0.05$) and one-way analysis of variance (ANOVA, $p < 0.001$) are employed to analyze the results [10, 11]. And the Dunnett test ($p < 0.05$) is done for the verification after test [12, 13].

3. Results

3.1. Surface roughness and strength

The surface roughness (arithmetic mean deviation of the profile, Ra) is examined by the laser surface roughness tester to investigate the effect of surface roughness on the degree of GBR and the biocompatibility between the Ti membrane surface and the gingival tissue. The surface-treated Ti barrier membranes are reported to replace the deficient and infected alveolar bone and provide the sufficient space underneath the membrane. The




Fig. 3. A comparison of surface roughness (Ra) in various surface-treated Ti membranes.

experimental results are shown in Figs. 1-3, indicating that the surface roughness (Ra) decreases as the treatment is changed from machining ($0.37 \pm 0.09 \mu\text{m}$), TiN coating ($0.22 \pm 0.09 \mu\text{m}$), electropolishing ($0.20 \pm 0.03 \mu\text{m}$), to anodizing ($0.15 \pm 0.03 \mu\text{m}$). The smoothest surface is observed for the anodized specimens.

The uniaxial tensile tests of the Ti membranes are performed. The specimens, surface-treated with electropol-




Fig. 4. Load-extension curves for the Ti membranes that are surface-treated: (a) machining, (b) TiN coating, (c) electropolishing, and (d) anodizing, respectively.

ishing or anodizing, exhibit better performance of deformation relatively at a low load as compared to the specimens treated with machining or TiN coating, as shown in Fig. 4. Although it is hard to compare the ductility (strain at break, %) due to different scale of x-axis, the ductility increased from machining (6.26 %), TiN coating (7.29 %), anodizing (16.82 %), to electropolishing (24.50 %). The ductility is deteriorated after machining and TiN coating as compared to electropolishing probably due to accumulation of residual stress. It is concluded that the electropolished specimen is likely to deform due to excellent ductility.

3.2. Biocompatibility

3.2.1. Cell adhesion and proliferation

The L-929 cell adhesion results are shown in Fig. 5, suggesting that no dramatic difference is detected. However, the electropolished specimen exhibits the lowest adhesion. The results are significant with a 95 % confidence level ($p > 0.05$). The cell proliferation results of L-929 cell are shown in Fig. 6, indicating that all the results are very similar and the 7-day proliferation is twice larger than the 3-day proliferation. The results are




Fig. 5. L-929 cell adhesion results of surface-treated Ti membranes.

significant with a 95 % confidence level ($p > 0.05$).

3.2.2. Osteoblast-like cell (MG-63) adhesion and proliferation

The results are exactly the same as those of L-929 cell adhesion and proliferation, as displayed in Figs. 7 and 8. The 7-day proliferation, as shown in Fig. 8,




Fig. 6. L-929 cell proliferation test results of surface-treated Ti membranes.




Fig. 7. MG-63 cell adhesion results of surface-treated Ti membranes.




Fig. 8. MG-63 cell proliferation test results of surface-treated Ti membranes.




Fig. 9. Cell viability of surface-treated Ti membranes.

increases 2.5 times larger than the 3-day proliferation. The results are significant with a 95 % confidence level ($p > 0.05$).

3.2.3. L-929 Cell cytotoxicity

All the results reveal that the test (Fig. 9) shows no evidence of causing cell lysis or toxicity. The results are significant with a 95 % confidence level ($p > 0.05$).

4. Discussion

GBR using barrier membranes is useful in bone augmentation. Commonly used polytetrafluoroethylene (PTFE) and Ti are the typical non-resorbable membranes. Although they need a secondary surgery to remove the membrane, the ingrowth of osteogenetic cells is not disturbed by competing non-osteogenetic soft tissue cells due to the superior stability of barrier membrane. They are likely to create and maintain sufficient space underneath the membrane for up to 6 months [1, 4, 5].

Immediate implant placement after tooth extraction may lead to early implant failure caused by insufficient mature bone mass and quality to adequately support the implant, and difficulty in obtaining complete coverage of the extraction socket by bone and soft tissue. Early membrane exposure by epithelial dehiscence may result in membrane exposure and infection, which can jeopardize bone augmentation. Kfir *et al.* [14] suggested that no apparent infection or problem is reported even though early Ti membrane exposure occurs between 2 and 6 weeks in 7 patients (47 %) out of 15 patients. Immediate bone augmentation using the Ti membrane after infected tooth extraction is found to be feasible and

safe. To prevent the early exposure, early bone augmentation after extraction of infected teeth and extensive bone loss is successfully achieved by using a Ti barrier membrane accompanied by socket filling with autologous platelet-rich fibrin (PRF) or bone graft material [15, 16].

The Ti membranes are more difficult to insert than flexible membranes due to sharp edges [4]. If the sharp edges of the Ti membrane cut through the mucosa, the bone augmentation should be damaged due to the membrane exposure and the related inflammation. In onlay grafts, it is essential to insert the membrane carefully to avoid the crumpling, when the surgeon is trying to bend it around two axes. Bending it around one axis is tolerable and the smooth surface is preserved. Careful soft tissue management is important to reduce the number of exposed membranes.

Experimental results revealed that surface roughness of the membrane was attributed to the surface treatments. The roughness decreased from 0.37 μm to 0.15 μm (machining (0.37 μm), TiN coating (0.22 μm), electropolishing (0.20 μm), anodizing (0.15 μm)). Strength and ductility decreased and increased with decreasing the surface roughness, respectively. In the present study, the electropolished specimen is determined to be the best membrane due to excellent ductility (24.50 %). A cytotoxicity test determines whether a product will have a toxic effect on living cells. Under the conditions of this study, the test extract showed no evidence of causing cell lysis or toxicity (Fig. 9) regardless of surface roughness, suggesting that the surface-treated specimens are suitable for the barrier membrane due to the absence of cytotoxicity. Cell adhesion results of L-929 and MG-63 (Figs. 5 and 7) indicated that the machined Ti membrane exhibited the highest cell adhesion while the electropolished membrane was determined to be the best membrane for the L-929 cell proliferation after 7 days. However, no appreciable difference in MG-63 cell proliferation between the surface-treated membranes was noticed.

In general, Ti membrane needs a secondary surgery to remove the barrier after sufficient bone formation underneath the Ti membrane. However, early exposure or dislocation of the membrane is possible especially at the interface between the Ti membrane and the incision, resulting in prevention of new bone growth caused by bacterial contamination (infection, contact between saliva and the grafted material). In that case, the electropolished Ti membrane is believed to be the best membrane due to excellent L-929 cell proliferation, as

Table 1
Tailored properties of surface-treated Ti membranes

Advantages	Surface-treatment condition	
Flexibility	Electropolishing, Anodizing	Easy to deform
Cell adhesion	Machining, TiN coating	MG-63 (osteoblast)
Cell proliferation	All	MG-63
Cell adhesion	Machining, Anodizing	L-929 (fibroblast)
Cell proliferation	Electropolishing, Anodizing	L-929

listed in Table 1. In the present study, it can be concluded that the electropolished Ti membrane is likely to be the best membrane due to tailored flexibility and excellent L-929 cell proliferation.

5. Conclusion

The Ti barrier membranes are prepared to evaluate the effect of surface-treatments, such as machining, electropolishing, anodizing, and electropolishing + TiN coating, on guided bone regeneration and flexibility of the membranes. The surface roughness (R_a) of the membrane decreases from machining (0.37 μm) to anodizing (0.15 μm). Strength and ductility decreased and increased with decreasing the surface roughness, respectively. No cytotoxicity causing cell lysis is observed regardless of the surface-treatments. Cell proliferation results of L-929 and MG-63 show that the electropolished membrane is the best membrane for the L-929 cell proliferation after 7 days. However, no appreciable difference in MG-63 cell proliferation between the surface-treated membranes was detected. It can be concluded that the surface-treated Ti specimens are suitable under the relevant circumstances that can arise from this experiment.

References

- [1] C.N. Elias, J.H.C. Lima, R. Valiev and M.A. Meyers, "Biomedical applications of titanium and its alloys", JOM, 46 (March 2008).
- [2] Y. Okazaki, "A new Ti-15Zr-4Nb-4Ta alloy for medical applications", Curr. Opinion Solid State Mater. Sci. 5 (2001) 45.
- [3] C.E. Wen, Y. Yamada, K. Shimojima, Y. Chino, T. Asahina and M. Mabuchi, "Processing and mechanical properties of autogenous titanium implant materials", J. Mater. Sci.: Mater. Med. 13 (2002) 397.
- [4] F. Watzinger, J. Luksch, W. Milleli, C. Schopper, J. Neugebauer, D. Moser and R. Ewers, "Guided bone regeneration with titanium membranes: A clinical study", British J. Oral. Maxillofac. Surg. 38 (2000) 312.
- [5] E. Kfir, V. Kfir and E. Kaluski, "Immediate bone aug-

- mentation after infected tooth extraction using titanium membranes", *J. Oral. Implantol.* 33 (2007) 133.
- [6] B.S. McAllister and K. Haghighat, "Bone augmentation techniques", *J. Periodontol.* 78 (2007) 377.
- [7] M. Hong, Y.H. Kim, D. Ganbat, D. Kim, C. Bae and D.S. Oh, "Capillary action: Enrichment of retention and habitation of cells via micro-channeled scaffolds for massive bone defect regeneration", *J. Mater. Sci.: Mater. Med.* 25 (2014) 1991.
- [8] D.S. Oh, Y.H. Kim, D. Ganbat, M. Han, P. Lim, J. Back, F.Y. Lee and H. Tawfeek, "Bone marrow absorption and retention properties of engineered scaffolds with micro-channels and nano-pores for tissue engineering: A proof of concept", *Ceram. Intl.* 39 (2013) 8401.
- [9] N.S. Turker, A.Y. Ozer, B. Kutlu, R. Nohutcu, A. Sungur, H. Bilgili, M. Ekizoglu and M. Ozalp, "The effect of gamma radiation sterilization on dental biomaterials", *Tissue Eng. Regen. Med.* 11 (2014) 341.
- [10] J. Kim and H. Kim, "Rat defect models for bone grafts and tissue engineered bone constructs", *Tissue Eng. Regen. Med.* 10 (2013) 310.
- [11] J. Kim, D.Y. Lee, T. Kim, Y. Song and N. Cho, "Biocompatibility of hyaluronic acid hydrogels prepared by porous hyaluronic acid microbeads", *Met. Mater. Intl.* 20 (2014) 555.
- [12] K. Chae, H. Choi, J. Ahn, Y. Song and D.Y. Lee, "Application of taguchi method and orthogonal arrays for characterization of corrosion rate of IrO₂-RuO₂ film", *J. Mater. Sci.* 37 (2002) 3515.
- [13] Y. Song, Y. Nam, K. Lee and D.Y. Lee, "Microstructure and chemical properties of c-BN films prepared by ME-ARE", *Mater. Sci. Forum* 449-452 (2004) 465.
- [14] E. Kfir, V. Kfir and E. Kaluski, "Immediate bone augmentation after infected tooth extraction using titanium membranes," *J. Oral. Implantol.* 33 (2007) 133.
- [15] J. Torres, F. Tamimi, M.H. Alkhraisat, A. Manchon, R. Linares, J.C. Prados-Frutos, G. Hernandez and E.L. Cabarcos, "Platelet-rich plasma may prevent titanium-mesh exposure in alveolar ridge augmentation with anorganic bovine bone," *J. Clin. Periodontol.* 37 (2010) 943.
- [16] W. Zhong, G. Ma, Y. Wang, R. Tamamura and J. Xiao, "Augmentation of peri-implant bone defects with different bone grafts and guided bone regeneration: a pilot experimental study in the dog", *J. Hard. Tissue Bio.* 15 (2006) 82.